



Optomagnetic Sequence-Specific Detection of Dengue Target DNA

Minero, Gabriel Khose Antonio; Nogueira, Catarina; Rizzi, Giovanni; Tian, Bo; Fock, Jeppe; Dolonato, Marco; Strömberg, Mattias ; Hansen, Mikkel Foug

Publication date:
2017

Document Version
Peer reviewed version

[Link back to DTU Orbit](#)

Citation (APA):

Minero, G. K. A., Nogueira, C., Rizzi, G., Tian, B., Fock, J., Dolonato, M., Strömberg, M., & Hansen, M. F. (2017). *Optomagnetic Sequence-Specific Detection of Dengue Target DNA*. Abstract from XIX Euroanalysis 2017, Stockholm, Sweden.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Optomagnetic Sequence-Specific Detection of Dengue Target DNA

Gabriel Antonio S. Minero¹, Catarina Nogueira², Giovanni Rizzi¹, Bo Tian³, Jeppe Fock¹,
Marco Dolonato², Mattias Strömberg³, and Mikkel F. Hansen¹

¹ Department of Micro- and Nanotechnology, Building 345B, DTU Nanotech, Technical University of Denmark, DK-2800 Kongens Lyngby, Denmark

² BluSense Diagnostics ApS, Symbion Bioscience Park, Fruebjergvej 3, DK-2100 Copenhagen, Denmark

³ Division of Solid State Physics, Department of Engineering Sciences, Ångström Laboratory, Uppsala University, Lägerhyddsvägen 1, SE-751 21 Uppsala, Sweden

Optomagnetic (OM) measurements, probing changes in the rotation dynamics of magnetic nanoparticles (MNPs) in response to an oscillating magnetic field, were recently used to detect products of loop-mediated isothermal amplification (LAMP) originating from Newcastle disease virus under ideal, clean lab conditions with a limit of detection (LOD) of 10 aM for 30 min assay time¹. There, a fraction of the forward inner primers was biotinylated, and the end-point OM readout was carried out after mixing the sample with streptavidin MNPs. Here, we report on the first real-time OM measurements during LAMP of Dengue target DNA under non-ideal lab conditions where the reaction and detection took place repeatedly in a single lab. Using an alternative detection scheme, we discriminated between spurious amplification products from previous reactions and specific amplification products from the actual sample. In this reaction, MNPs functionalized with capture probes, targeting the loops of specific LAMP amplicons (Fig 1A), were premixed with the sample. Capture probes bound to spurious LAMP products containing mutations had a 20°C lower melting point (Fig. 1C) allowing for discrimination between the two types of products at 67°C (Fig. 1A-B). In LAMP, strand displacement and self-priming lead to long single stranded tracts containing repeats of the loop sequences². Specific binding of MNPs to these resulted in MNP clustering (signal at $f < 20$ Hz) and a depletion of free MNPs (signal at $f \approx 300$ Hz), see Fig. 1A. Fig. 1D shows the decrease of the single MNP signal vs. Dengue target concentration observed at 67°C. By this, we found an LOD of about 100 fM after 20 min of amplification.




Fig. 1: Sequence-specific validation of Dengue amplicons in LAMP detected in the OM signal. (A) Capture probe (green) binding to the complementary DNA sequence of specific Dengue amplicons in the LAMP elongation phase. (B) Capture probe do not bind to spurious amplicons with mutations (blue). Curves show the optomagnetic signal at different times. (C) Melting of DNA bridges between amplicons and MNPs via probing the median hydrodynamic diameter of the MNPs. (D) Dose-dependent analysis of Dengue target amplification at a LAMP threshold time of 20 min (67°C). Two blue lines represent 3 σ rule of thumb for calculation of LOD.

1 B. Tian et al., *ACS Sensors*, 2016, **1**, 1228–1234.

2 T. Notomi, Y. Mori, N. Tomita and H. Kanda, *J. Microbiol.*, 2015, **53**, 1–5.